



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Thayer Police Department
Serial Number: SD2301
Manufacturer: Guth
Model Number: 10-4D

CALIBRATION RESULTS


<u>Reference</u> <u>Temperature</u>	<u>Simulator</u> <u>Temperature</u>
33.98	34.0

This calibration was performed with
NIST-Traceable Thermometer SN: 304454

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 04/03/2015

COPY OF CALIBRATION STICKER

This simulator has been calibrated according to DHSS specifications	
	SIMULATOR SERIAL NO.: <u>SD2301</u>
	EXPIRATION DATE: <u>04/03/2016</u>
	DATE OF CALIBRATION: <u>04/03/2015</u>
	NIST REF. THEM. SERIAL NO.: <u>304454</u>
	AVERAGE SIM. TEMP: <u>33.98 C</u>
	ANALYST INITIALS: <u>BML</u>